

ANNEXURE A

SCHEDULE OF ACCREDITATION

DIMENSIONAL METROLOGY

Laboratory Accreditation Number: **CAL-2 006 (ISO/IEC 17025:2017)**

<p>Permanent Address of Laboratory Namibian Standards Institution Metrology Department 205 Gold Street, Prosperita Windhoek Namibia</p> <p>Postal Address P O Box 26364 Windhoek Namibia</p> <p>Tel : +264 61 386 470/481 Cell : +264 81 261 3694 Fax : +264 61 386 477 Email : matalis@nsi.com.na</p>		<p>Technical Signatories : Ms L Katanga (All items) Mr S Sankwasa (All items)</p> <p>Nominated Representative : Mr S Matali</p> <p>Issue No : 01 Date of Issue : 08 January 2025 Expiry Date : 27 February 2028</p>		
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)
			At NSI	
6.1	Hand Instruments			
6.1.1	External Micrometers	Internal: MTPI Reference: SADCAS TR20	0 to 25 mm 25 to 75 mm	0,002 mm 0,012 mm
			At NSI	
6.1.4	Caliper (Vernier, dial and electronic)	Internal: MTPI Reference: SADCAS TR20	0 to 150 mm 150 to 500 mm	0,02 mm 0,04 mm
			At NSI	
6.1.8	Dial Indicator	Internal: MTPI Reference: SADCAS TR20	0 to 25 mm	0,002 mm

Original date of accreditation: 08 January 2025

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%.



Pinkie J Malebe
SADCAS Technical Manager