

ANNEXURE A

SCHEDULE OF ACCREDITATION

MASS METROLOGY

Laboratory Accreditation Number: **CAL-8 002 (ISO/IEC 17025:2017)**

<p>Permanent Address of Laboratory Namibian Standards Institution Metrology Department 205 Gold Street, Prosperita Windhoek Namibia</p> <p>Postal Address P O Box 26364 Windhoek Namibia</p> <p>Tel : +264 61 386 470/481 Cell : +264 81 261 3694 Fax : +264 61 386 477 Email : matalis@nsi.com.na</p>		<p>Technical Signatories : Mr S Matali (All items) Mr S S Sankwasa (All items) Ms L Katanga (All items)</p> <p>Nominated Representative : Mr S Matali</p> <p>Issue No : 02 Date of Issue : 20 August 2025 Expiry Date : 27 February 2028</p>		
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)
1	Mass pieces	Internal: <i>MPTI 002</i> ; Reference: OIML R111-1	1 mg to 5 g 10 g 20 g to 50 g 100 g 200 g 500 g 1 kg to 2 kg 5 kg 10 kg 20 kg	At NSI 0.02 mg 0.03 mg 0.04 mg 0.1 mg 0.2 mg 0.001 g 0.002 g 0.005 g 0.03 g 0.03 g

Original date of accreditation: 20 February 2013

Page 1 of 2

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%.

ANNEXURE A

Facility Accreditation No: CAL-8 002 (ISO/IEC 17025:2017)

Issue No: 02

Date of Issue: 20 August 2025

Date of Expiry: 27 February 2028

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)
			At NSI and On-site	
2	Weighing Instruments <ul style="list-style-type: none"> Digital self- indicating 	Internal: MPTI 003 Reference: OIML R76, EURAMET cg 18	0 g to 50 g 50 g to 200 g 200 g to 3000 g 3.0 kg to 1500 kg	0.00002 g 0.0006 g 0.00015 % + 0.00 g 0.001 %

Original date of accreditation: 20 February 2013

Page 2 of 2

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%.

Pinkie J Malebe
SADCAS Technical Manager