



ANNEXURE A

SCHEDULE OF ACCREDITATION

DIMENSIONAL METROLOGY

Laboratory Accreditation Number: **CAL-2 006 (ISO/IEC 17025:2017)**

Permanent Address of Laboratory Namibian Standards Institution Metrology Department 205 Gold Street, Prosperita Windhoek Namibia		Technical Signatories : Ms L Katanga (All items) Mr S Sankwasa (All items)		
Postal Address P O Box 26364 Windhoek Namibia		Nominated Representative : Mr S Matali		
Tel : +264 61 386 470/481 Cell : +264 81 261 3694 Fax : +264 61 386 477 Email : matalis@nsi.com.na		Issue No : 02 Date of Issue : 20 August 2025 Expiry Date : 27 February 2028		
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)
			At NSI	
6.1	Hand Instruments			
6.1.1	External Micrometers	Internal: MTPI 022 Reference: SADCAS TR20	0 to 25 mm	0,002 mm
			At NSI	
6.1.4	Caliper (vernier, dial and electronic)	Internal: MTPI 023 Reference: SADCAS TR20	0 to 150 mm 150 to 500 mm	0,02 mm 0,04 mm
			At NSI	
6.1.8	Dial Indicator	Internal: MTPI 021 Reference: SADCAS TR20	0 to 25 mm	0,002 mm

Original date of accreditation: 08 January 2025

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%.

Pinkie J Malebe
SADCAS Technical Manager