

NAMS 14562:2024

First Edition

EN 14562:2006

NAMIBIAN STANDARD

Chemical disinfectants and antiseptics - Quantitative carrier test for the evaluation of fungicidal or yeasticidal activity for instruments used in the medical area - Test method and requirements

This Namibian standard is the identical adoption of EN 14562:2006 and is adopted with the permission of the European Committee for Standardization

Published by the Namibian Standards Institution (NSI)
Established by section 2 of the Standards Act, 2005 (Act No 18 of 2005)
37 Feld Street
P.O. Box 26364 Windhoek, Namibia
Tel +264-61386400, Fax +264-61-386454
Website: www.nsi.com.na
© NSI



Licensed by NSI to NSI for internal use only
DOWNLOADED:24/06/2026
Single-user licence only, copying and networking prohibited.

NAMS 14562:2024

First Edition

EN 14562:2006

National foreword

This Namibian Standard (NAMS) is identical to EN 14562:2006, and was approved for adoptions by the Namibian Standards Institution CEO.

Namibian standards are developed based on NSI Standards development procedures in accordance with the rules given in the International Organisation for Standardisation/ International Electrotechnical Commission (ISO/IEC) Directives 1, ISO/IEC Guide 21-1 Adoption of international standards as regional or national standards and WTO – TBT World Trade Organisation code of Good Practice (which is published as Annex 3 in the TBT Agreement)

The NSI Technical Committee responsible for this standard is NSI TC 14, Chemicals.

This NAMS 14562:2024 EN 14562:2006 was published in May 2024.

English Version

Chemical disinfectants and antiseptics - Quantitative carrier test
for the evaluation of fungicidal or yeasticidal activity for
instruments used in the medical area - Test method and
requirements (phase 2, step 2)

Désinfectants et antiseptiques chimiques - Essai quantitatif
de porte germe pour l'évaluation de l'activité fongicide ou
levuricide pour instruments utilisés en médecine humaine -
Méthode d'essai et prescriptions (phase 2, étape 2)

Chemische Desinfektionsmittel und Antiseptika -
Quantitativer Keimträgerversuch zur Prüfung der fungiziden
oder levuroziden Wirkung für Instrumente im
humanmedizinischen Bereich - Prüfverfahren und
Anforderungen (Phase 2, Stufe 2)

This European Standard was approved by CEN on 29 August 2005.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

page

Foreword	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Requirements	6
5 Test method.....	6
5.1 Principle.....	6
5.2 Materials and reagents	7
5.3 Apparatus and glassware.....	9
5.5 Procedure for assessing the fungicidal / yeasticidal activity of the product	14
5.6 Experimental data and calculation	18
5.7 Verification of methodology.....	23
5.8 Expression of results and precision.....	24
5.9 Interpretation of results – conclusion	25
5.10 Test report	26
Annex A (informative) Referenced strains in national collections.....	28
Annex B (informative) Suitable neutralizers	29
Annex C (informative) Graphical representations of the test method	31
Annex D (informative) Example of a typical test report	33
Annex E (informative) Information on the application and interpretation of European Standards on chemical disinfectants and antiseptics.....	37
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC.....	39
Bibliography.....	40

Foreword

This European Standard (EN 14562:2006) has been prepared by Technical Committee CEN/TC 216 “Chemical disinfectants and antiseptics”, the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2006, and conflicting national standards shall be withdrawn at the latest by November 2006.

Other methods to evaluate the efficacy of chemical disinfectants and antiseptics for different applications in the medical field are in preparation.

A collaborative trial will be undertaken to provide a precision annex to this standard.

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this standard.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Introduction

This European Standard specifies a carrier test for establishing whether a chemical disinfectant for use on instruments (surgical instruments, anaesthesia material, endoscopes etc.) has a fungicidal or yeasticidal activity in the fields described in the scope.

The laboratory test closely simulates practical conditions of application including pre-drying fungi on a carrier, contact time, temperature, test organisms and interfering substances i.e. conditions which may influence the action of chemical disinfectants in practical situations.

The obligatory conditions are intended to cover general purposes and to allow reference between laboratories and product types. Each utilization concentration of the chemical disinfectant found by this test corresponds to defined experimental conditions. However, for some applications the recommendations and/or instructions of use of a product may differ and therefore additional test conditions need to be used.